



NOTES

1. Thickness changed in Rev 1 to account for actual thickness of PFS07003 as well as light-weighting in that part. Original thickness for PFS07004 was 0.257 to properly counterweight PFS07003 with a nominal thickness of 0.063.

**THE OBSERVATORIES
OF THE CARNEGIE INSTITUTION OF WASHINGTON**
813 Santa Barbara Street
Pasadena, CA 91101

MATERIAL		UNLESS OTHERWISE SPECIFIED		
Brass		FRACTIONAL +/- 1/32	DECIMAL .XX +/- 0.01	ANGULAR +/- 0.1 DEG
EST WEIGHT		.XXX +/- 0.002 GOAL .XXX +/- 0.005 REQD		
0.79 lb		BREAK SHARP EDGES MACHINED SURFACES 125° ALL DIMENSIONS IN INCHES		
SCALE		1:1		Planet Finder Spectrograph
DWN	CRANE	7/20/2007		Hartmann Mask Counterweight
APVD				
BY		DATE		SHEET A
Bill of Material		DRAWING NUMBER PFS07004		SHT 1 OF 1
				REV 1

1	1	PFS07004
ITEM	QTY	DESCRIPTION
Bill of Material		

2

1

2

1